






Educational Grant Resources: Endosee Training Materials








This worksheet provides a list of resources available to support hysteroscopy workshops along with product numbers. The completed form must be submitted as an attachment to your CooperSurgical Grant Application.

Requests for Endosee handsets for use in workshops require organization to agree to have materials accompanied by CooperSurgical, Inc. representative to set up and maintain safe operation of equipment (see application). If you would like to discuss availability of options such as papaya or bell pepper stations, indicate interest on accompanying application and we will contact you.

For questions regarding materials or orders, please contact mededgrants@coopersurgical.com.

Name of Workshop(s) where equipment will be used: _____

Name	Description	Product #	Requested #
Endosee® Advance Display Module		ES-9000	
Endosee® Advance Cannula		ESPX5	
Biospy Forceps	 <p>EP-BSPY Biospy forceps</p>	ES-BSPY	
Spoon Forceps	 <p>ES-LNGR Spoon forceps, long serrated</p>	ES-LNGR	
Foreign Body Grasper	 <p>ES-FBGR Foreign body grasper, fenestrated</p>	ES-FBGR	

Scissors	 <p>ES-SCIS Scissors, single action</p>	Endosee® Advance working channel instrument: Single action scissors Typically, 1-2 per handset is sufficient per request.	ES-SCIS	
50 ml Syringe		Typically 1-2 per handset	ESSYR	
Standard Bore Extension Tubing		Typically 1-2 per handset	ESTUB	
Table-Top Pelvic Model (Lazarus)		Table-top pelvic model. Allows learner to visualize uterine anatomical markers using Endosee	LAZARUS	
Hysteroscopy Coding Guide		Coding Guide Reference tool	Coding Guide Reference tool	
Endosee Gen I Handset		Limited availability, Endosee Gen I does not function with working channel instruments	ES-8000	
Endosee Gen I Cannula		Workshop only: These cannulas can be re-used. Typically, 1-2 cannulas per handset is sufficient per request.	ESDX5	